




TPP “easy read” Measuring Device Measuring Accuracy



Measuring accuracy

The TPP “easy read” measuring gauge has an accuracy of ± 0.01 mm.

Measurement Point	Volume on Scale μl	Maximum Deviation from Nominal mm
1	5.0	± 0.05
2	3.5	± 0.05
3	0.7	± 0.05
4	0.3	± 0.05
5	0.0	± 0.05

Disclaimer

TPP products are intended for Research Use Only (RUO) and are not approved for clinical, diagnostic, or in vitro fertilization (IVF) applications. The full Terms & Conditions, including limitations of warranty and liability, intended use, and reseller obligations, are available at:

https://www.tpp.ch/page/qualitaets_sicherung/index.php

Distributors who purchase and distribute TPP products acknowledge and agree to these Terms & Conditions and the associated disclaimer.